

Declaration of Conformity

Cochlear Limited of 1 University Avenue, Macquarie University, NSW 2109, Australia, declares under our sole responsibility that the following product, to which this declaration relates, is in conformity with the listed standards:

Product name: 1. Nucleus® CP810 Processing Unit
 2. Nucleus® CR110 Remote Assistant
 Model number: 1. CP810
 2. CR110

Standards:


Essential Requirements	Standards	Report #	
		CP810	CR110
Health & Safety	EN 60601-1:2006 + Corr.1:2010 (IEC 60601-1:2005+Corr.1:2006+Corr.2:2007)	384326	
EMC	EN 60601-1-2:2007 (IEC 60601-1-2:2007)	315564	315564
	EN 301 489-1 v2.2.0	D1246984	D1246985
	EN 301 489-17 v3.2.0	D1246984	D1246985
Radio	EN 300 328 v2.1.1	D1246986	D1246987

We hereby declare that all essential radio test suites have been carried out and that the above named products are in conformity to all the essential requirements of Directive 2014/53/EU.

The conformity assessment procedure referred to in Article 17 and detailed in ANNEX II of Directive 2014/53/EU has been followed. The technical documentation relevant to the above equipment is kept at our address listed above.

I declare that Cochlear Limited is exclusively responsible for this Declaration of Conformity.

Authorised Signatory:



Sydney, Australia; 03-Nov-2017

Robert Gay, Senior Regulatory Affairs Manager

Place and Date