

DECLARATION OF CONFORMITY

Cochlear Limited of 1 University Avenue, Macquarie University, NSW 2109, Australia, declares under our sole responsibility that the following product, to which this declaration relates, is in conformity with the listed standards:

Product name: 1. Nucleus® 7 Processing Unit
 2. Cochlear™ Remote Control

Model number: 1. CP1000
 2. CR310

Standards:

| Essential Requirements | Standards | Report # |
|------------------------|---|--|
| Health & Safety | EN 60601-1:2006+Corr.1:2010+A11:2011+A12:2014 (IEC 60601-1:2005+Corr.1:2006+Corr.2:2007+A1:2012) | D893055 |
| EMC | EN 60601-1-2:2015 (IEC 60601-1-2:2014) EN 301 489-1 v2.1.1 EN 301 489-17 v3.1.1 | D818571 D827044 D953916, D1238655 D953815, D1238656 |
| Radio | EN 300 328 v2.1.1 | D971200, D1204041, D1226413; D971395, D1226414 |

We hereby declare that all essential radio test suites have been carried out and that the above named products are in conformity to all the essential requirements of Directive 2014/53/EU.

The conformity assessment procedure referred to in Article 17 and detailed in ANNEX III of Directive 2014/53/EU has been followed with the involvement of the following Notified Body:

- DEKRA Testing and Certification, S.A.U., Parque Tecnológico de Andalucía, C/ Severo Ochoa nº 2, 29590 Campanillas – Malaga, Spain.
- Notified Body No.: 1909.

The technical documentation relevant to the above equipment is kept at our address listed above.



 Mark Smyth, Head of Global Regulatory Affairs

Sydney, Australia; 09-June-2017

 Place and Date