



Cochlear[™] Osia[®] Surgical Instruments Sterilisation Reprocessing Guide

Symbols used in this guide



Note

Important information or advice.



Caution (no harm)

Special care to be taken to ensure safety and effectiveness. Could cause damage to equipment.



Warning (harmful)

Potential safety hazards and serious adverse reactions. Could cause harm to person.

Contents

	Symbols used in this guide2	
In	troduction4	
	Bone bed indicator 17 mm	
	Limitations on reprocessing5	
	Instrument material compatibility5	
Re	eprocessing instructions	,
	Point of use	,
	Containment and transportation6)
	Preparation for cleaning	
	Cleaning and disinfecting	,
	Automated cleaning and disinfecting procedure 8 $$	
	Drying	1
	Maintenance10	
	Inspection and function testing10	
	Packaging	
	Sterilisation	
	Storage	
	Symbols	
Αŗ	pendix 1 – Validation information14	
	Test items14	
	Automated cleaning14	
	Thermal disinfection	1
	Steam sterilisation	
	Drying	
	Repeat reprocessing	ì
	Standards	į

Introduction

This guideline is intended for staff involved in sterilisation reprocessing of the Bone bed indicator 17 mm. The Bone bed indicator 17 mm is intended to be used throughout the surgery to correctly position and attach the OSI200 or OSI300 Implant. For information on instruments used during the surgery, refer to the Physician's Guide.

Bone bed indicator 17 mm



P1469690

Bone bed indicator 17 mm

- Delivered in two parts (body and pin) that have to be combined before use.
- Parts do not lock.

The sterilisation department or reprocessing centre at your hospital or clinic is responsible for surgical instrument sterility. They should:

- Use device and product-specific validated procedures for cleaning, disinfecting and sterilisation.
- Use washer-disinfectors and sterilisers that are maintained and checked regularly.
- Make sure recommended parameters are applied for each cycle.

Cochlear has validated the instructions in this guide for preparing a surgical instrument for re-use. Staff at the hospital or the clinic are responsible for ensuring that reprocessing achieves the desired result—as performed using equipment, materials and staff in the reprocessing centre or sterilisation department. This requires validation and routine monitoring of the process. Any deviation from the instructions provided should be properly evaluated for effectiveness and potential adverse consequences.

Marning

- The Cochlear Bone bed indicator 17 mm is supplied non-sterile and must be cleaned, disinfected and sterilised before use.
- · Instruments must not contact other instruments during cleaning and disinfecting.
- · Damaged instruments must not be used.
- · Do not use:
 - Combined cleaning-disinfection solutions
 - Dry heat, radiation, formaldehyde, ethylene oxide or plasma sterilisation
 - Instrument oils for instrument maintenance no maintenance required
- Follow the legal provisions for your country and the hygiene instructions of your hospital or clinic. This particularly applies to different guidelines regarding the inactivation of prions.

Limitations on reprocessing

Repeated processing has a minimal effect on these instruments. The instruments have been validated for 25 cycles of reprocessing as instructed in this guide. End-of-life is normally determined by visible manifestation of wear and damage. See "Inspection and function testing" on page 10.

⚠ Caution

Do not use metal brushes or steel wool for cleaning.

Instrument material compatibility

- To avoid corrosion, do not process Cochlear metal instruments with instruments that have aluminium, brass, copper or chrome-plated parts.
- Do not use cleaning detergents with the following ingredients:
 - Organic, mineral, and oxidising acid. The minimum allowed pH-value is 5.5.
 - Halogens (for example chlorine, iodine, bromine).
 - Aromatic, halogenated hydrocarbons.
- Do not expose instruments to temperatures higher than 142 °C (288 °F).

Reprocessing instructions

The following reprocessing instructions are for instruments used with patients who in the general population represent no identified risk of Transmissible Spongiform Encephalopathies (TSEs) transmission and where:

a. The instruments have not been exposed to tissue that are known to have high-infectivity for Creutzfeldt-Jakob disease (CJD) (e.g. dura)

or

b. Where the instruments have been exposed to tissue known to have high-infectivity for CJD (e.g. dura).



- The reprocessing should be performed by suitably trained staff using well-maintained equipment in a facility that meets the requirements of ISO 17665- 1, for equipment validation and routine control.
- For instruments used with patients who represent a definite or potential risk of TSE transmission, contaminated instruments should be placed immediately into the correct clinical waste container for disposal. Follow the legal provisions for your country and the hygiene instructions of your hospital or clinic.

Point of use

Throughout the surgical procedure, wipe blood and debris from instruments to prevent drying. Do not allow contaminated devices to dry before reprocessing.

Directly after use, remove coarse impurities from the instruments by wiping the articles using lint-free cloths dampened with distilled water and soak in distilled water until they are reprocessed.

Instruments should be reprocessed within one hour after use.

Containment and transportation

Follow the validated containment and transportation instructions of your hospital or clinic.

Preparation for cleaning


Rinse under water at a temperature that should not exceed 45 °C (113 °F) while brushing for a minimum of 1 minute.

A Caution

Do not use metal brushes or steel wool. For manual removal of impurities, use a soft brush only. For holes, use suitable interdental brushes.

Bone bed indicator 17 mm

- The Bone bed indicator 17 mm consists of a body and a pin
- Disassembling required if assembled



Cleaning and disinfecting

Marning

Do not use a manual procedure, even in conjunction with an ultrasonic bath. Manual procedures could cause damage to delicate instrument parts, resulting in prolonged surgery time.

Equipment: Washer-disinfector and cleaning detergent.

The washer-disinfector must have these properties:

- Approved efficiency (CE mark in Europe, FDA registration/clearance for USA).
- · Validated for EN ISO 15883.
- Approved program for thermal disinfection (A₀ value >3000).
- Suitable program for the instruments with sufficient rinsing steps.

The cleaning detergent must have the following properties:

- Suitable for cleaning stainless steel instruments.
- Compatible with the instruments (see "Instrument material compatibility" on page 5).



Cleaning validation as shown in "Appendix 1 – Validation information" on page 14, was undertaken using minimum detergent concentration levels. Always follow the instructions from the detergent manufacturer regarding concentration, temperature and soaking time.

Apply only the parameters that were applied during process validation in the hospital or clinic's reprocessing centre or sterilisation department.



For instruments in contact with patients with no identified risk of TSE transmission and where the instruments have been exposed to tissue known to have high infectivity for CJD (e.g. dura), fully immerse the instruments in a 1% solution of Neodisher® MediClean Forte at 55 °C (131 °F) and stir for 10 minutes prior to

Automated cleaning and disinfecting procedure

automated cleaning/disinfecting.

1. Transfer instruments into the washer-disinfector in a small parts basket. Position instruments to allow for drainage of water.



A Warning

Instruments must not contact other instruments during cleaning and disinfecting.

- 2. Start the validated program as described in Table 1 on page 9.
- 3. Execute the cycle release by checking that the process parameters were correctly applied.
- 4. Inspect instruments to ensure no residual moisture is present.
- 5. If required, use filtered, pressurised air to complete the drying process.

Cycle	Time	Minimum temperature	Detergent/water type
Pre-cleaning	2 minutes	Cold <40 °C (<104 °F)	Tap water
Detergent wash	2 minutes	Heated 40 °C – 55 °C (104 °F – 131 °F)	Enzymatic detergent OR Prion inactivating alkaline detergent*
Wash	5 minutes	Set point 55 °C (131 °F)	Neutral/non-enzymatic detergent
Rinse	2 minutes	Heated 50 °C – 60 °C (122 °F – 140 °F)	Hot tap water
Thermal disinfection	5 minutes	Heated 93 °C (200 °F)	Critical water [†]
Dry [‡]	10 minutes	Heated 110 °C (230 °F)	Not applicable

Table 1: Automated cycle parameters

* Warning

Use an alkaline detergent validated for prion inactivation (e.g. Neodisher® MediClean Forte) for instruments in contact with patients with no identified risk of TSE transmission and where the instruments have been exposed to tissue known to have high-infectivity for CJD (e.g. dura).

- † Water extensively treated to ensure removal of the microorganisms and the inorganic and organic material. Treatment is usually a multistep process that may include a carbon bed, softening, DI, and RO or distillation (maximum 10 germs/ml, maximum 0.25 endotoxin units/ml).
- If instruments are not dry after automated cleaning / disinfection cycle, filtered, pressurised air is recommended to complete the drying process.
 Follow the instructions of your hospital or clinic.



For validation information, see "Appendix 1 – Validation information" on page 14.

Drying

Inspect instruments to ensure no residual moisture is present. If required, use filtered, pressurised air to complete the drying process.

Hot air drying is not recommended, except for drying that is part of the abovementioned validated automatic washing-disinfecting cycle.

Maintenance



Marning

Do not use instrument oils for instrument maintenance-no maintenance required.

Inspection and function testing

Check all instruments after cleaning-disinfecting for corrosion, damaged surfaces and impurities.



Marning

Do not use damaged or worn instruments. Instruments that remain dirty must be cleaned and disinfected again.

For specific critical control points see Table 2 below.

Cochlear Osia surgical instruments	Specific critical control points Give particular attention to the following:
P1469690 Bone bed indicator 17 mm	Body: Internal surface of hole; Laser marking (UDI) Pin: Outer threads; Laser marking (UDI)

Table 2: Critical control points for surgical instrument inspection

Packaging

Pack the cleaned and disinfected instruments in sterilisation packaging that meets the following requirements:

- Compliant with EN ISO/ANSI AAMI ISO 11607.
- Sufficiently protects the instruments against physical damage.
- Maintains sterility of the instruments during handling and storage prior to use.
- Ensures the instruments are not in excessive contact with each other.
- Allows sufficient inner volume to avoid strain on the instruments.
- Sealed using a validated thermo-sealing process.

For more information see "Appendix 1 – Validation information" on page 14.

Sterilisation

Cochlear has developed and validated the sterilisation instructions in this guide for preparing a surgical instrument for re-use to comply with the requirements of ISO 17665-1.

 Transfer instruments into the steriliser avoiding known 'cold' locations, typically over the drain.

Use steam sterilisation as described:

- Dynamic air removal steam sterilisation cycle (prevacuum).
- Sterilisation parameters and instructions as specified by the steam steriliser
 manufacturer and your hospital or clinic, including those related to the quality of
 water supplied to the steam generator.
- Validated parameters for temperature and time for instruments used with patients with no identified risk of TSE transmission and where the instruments have not been exposed to tissue that are known to have high-infectivity for CJD (e.g. dura):

Minimum temperature	Minimum exposure time	Minimum drying time
132 °C (270 °F)	4 minutes	20 minutes
134 °C (273.2 °F) [†]	3 minutes	16 minutes

Use the following parameters for instruments used with patients with no identified risk of TSE transmission and where the instruments have been exposed to tissue that are known to have high-infectivity for CJD (e.g. dura):

Minimum temperature	Minimum exposure time	Minimum drying time
134 °C (273.2 °F)†	18 minutes*	16 minutes

- The construction and design of the surgical instruments have been verified to withstand 18 minutes exposure time at 134 °C (273.2 °F)
- † The cycles 134 °C (273.2 °F) for 3 minutes and 134 °C (273.2 °F) for 18 minutes are not applicable to U.S. healthcare facilities.



Marning

Do not use dry heat, radiation, formaldehyde, ethylene oxide or plasma sterilisation.

A Caution

Maximum sterilisation temperature of 138 °C (280 °F).

Storage

After sterilisation, store the instruments in sterilisation packaging in a dry and dust-free environment.

Symbols



Consult instructions for use



Catalogue number



Caution



Batch code



Date of manufacture



Non-sterile



Manufacturer



Caution: US law restricts this device to sale by, or on the order of, a physician



Authorised representative in the European Community



Medical device

Appendix 1 – Validation information

Test items

Samples of Cochlear's reusable surgical instruments with the most challenging features to clean, disinfect and sterilise were used in the validation tests.

Automated cleaning

Critical cleaning parameters were determined by rigorous automated cleaning methods developed using spore logarithmic reduction, total protein, haemoglobin and visual inspection criteria.

For device contamination, a blood soil containing 2M calcium chloride (for coagulation purposes) and bone meal (1 g/100 mL of test soil) was inoculated with a minimum population of 10⁴ CFU/mL of *G. stearothermophilus* and *E. faecium*.

The instruments were exposed to the test soil for a minimum of 15 minutes and allowed to dry for a minimum of 1 hour before cleaning and disinfecting. Any pre-cleaning involved rinsing the instrument under water (49 °C; 120.6 °F) while brushing with a Spectrum M-16 brush for a minimum of 1 minute.

After cleaning, test samples were visually inspected for any sign of remaining blood soil. Bioburden extractions were used to determine the number of spores, total protein and haemoglobin remaining on test samples. Comparison of data to positive controls and test protocol requirements determined if acceptance criteria were met.

Cleaning equipment				
Equipment	Washer/Disinfector	Steriliser		
Type	N/A	Steam; SG-120; AMSCO		
Manufacturer	Steris	AMSCO		
Serial Number	3603513001	0117594-02		
Model	GEN FH07-1XX	SG-120		

Table 3: Cleaning equipment used for cleaning validations using enzymatic detergent

Automated cleaning validations				
Cycle	Time	Validated settings	Detergent/water type	
Pre-cleaning	2 minutes*	Cold <40 °C (<104 °F) [†]	Tap water	
Detergent wash*	2 minutes*	Heated 48 °C (118 °F) [†]	Valsure Enzymatic Detergent 2mL/L	
Wash	5 minutes*	Set point 55 °C (131 °F)†	Valsure Neutral 2mL/L [†]	
Rinse	2 minutes*	Heated 50 °C – 60 °C (122 °F – 140 °F) [†]	Hot tap water	
Dry	10 minutes*	Heated <84.2 °C (183.5 °F)	Not applicable	

Table 4: Validation for automated cycles using enzymatic detergent

- * Validated exposure time required to achieve >3 log bioload reduction.
- † Validated exposure temperature required to achieve >3 log bioload reduction.

For instruments used with patients with no identified risk of TSE transmission and where the instruments have been exposed to tissue that are known to have high-infectivity for CJD (e.g. dura) the articles were soaked for in a 0.5% (5 mL/L) Neodisher Mediclean Forte solution at 55 °C whilst stirring for 10 minutes prior to the following automated cleaning cycle.

Cleaning equipment			
Equipment Washer/Disinfector Steriliser			
Туре	N/A	Steam; SG-120; AMSCO	
Manufacturer	Beli Med	AMSCO	
Serial Number	2005453	0117594-02	
Model	WD 290	SG-120	

Table 5: Cleaning equipment used for cleaning validations using alkaline detergent

Automated cleaning validations				
Phase	Recirculation time*	Temperature †	Detergent type and concentration (if applicable)	
Pre-wash	2 minutes *	Cold tap water	N/A	
Wash 1*	2 minutes *	43 °C (109 ° F) [†]	Neodisher Mediclean	
		Tap water	Forte 2 mL/L [‡]	
Wash 2	4 minutes *	55 °C (131 °F)†	Valsure Neutral	
		Tap water	2 mL/L	
Rinse	1 minute *	50 °C (122 °F)†	N/A	
		Tap water		

Table 6: Validation for automated cycles using alkaline detergent

- * Validated exposure time required to achieve >3 log bioload reduction.
- † Validated exposure temperature required to achieve >3 log bioload reduction.
- ‡ Neodisher® MediClean Forte, has been validated by the manufacturer for prion inactivation.



Testing validation was undertaken using minimum detergent concentration levels. Always follow the detergent manufacturer instructions regarding concentration levels when reprocessing instruments.

Thermal disinfection

Thermal disinfection parameters (5 minutes, 93 °C (199.4 °F) using critical water) were validated to demonstrate that Cochlear's reusable surgical instruments and selected worst-case thermocouple/temperature probe locations demonstrate an $A_0 \ge 600$. See "Table 3: Cleaning equipment used for cleaning validations using enzymatic detergent" on page 14 for details of the Washer/Disinfector used for thermal disinfection validation.

Steam sterilisation

Prevacuum (pressure pulse) sterilisation cycles were used for validation. Test articles were individually single-pouched in a 5.5×10 in. pouch (Cardinal Health self-sealed pouch CAT #92510 - 510(k) K153540) and placed on the edge in the steriliser.

Instruments were evaluated to a sterility assurance level (SAL) of \leq 10⁻⁶ using half-cycle studies and the biological indicator overkill method. Geobacillus stearothermophilus, ATCC #7953, was the indicator organism.

Sterilisation parameters				
Parameter	Setting 1			
Steriliser type	Prevacuum			
Preconditioning pulses 4 (Set point: 10 inHg for 1 minute)		inute)		
Temperature	132 °C (270 °F)	134°C (273.2°F)		
Full cycle time	4 minutes	3 minutes		

Table 7: Validation for sterilisation

Sterilisation equipment				
Manufacturer	Steris	Primus		
Model	LV-250	PSS8-A-MSSD		
Serial number	0305312-14 / 0305312-15 /	17730		
	0305412-25			

Table 8: Steam sterilisation validation equipment

Quality of steam was >97%.

Drying

Test articles were weighed pre-sterilisation and post-sterilisation to detect any residual moisture. Any residual moisture after drying in the cleaning/disinfecting cycle was removed by using pressurised filtered air before being prepared for sterilisation. All surfaces of test articles were inspected for visible moisture. Any visible moisture on pouch surfaces was noted.

Each test article was:

- Prepared as described in "Steam sterilisation" on page 17.
- Placed into the 'cold' spot of the steam steriliser, typically over the drain.
- Sterilised using validated full cycle set points with drying times in Table 9 below.

Parameter	Setting 1	Setting 2
Drying time	20 minutes	16 minutes

Table 9: Validation for drying

Repeat reprocessing

The instruments have been validated for 25 cycles of reprocessing as instructed in this guide.

Standards

Validation testing was performed using applicable standards.

For details please contact a Cochlear representative.

This page is intentionally left blank.

AU Cochlear Ltd (ABN 96 002 618 073)

1 University Avenue, Macquarie University, NSW 2109, Australia

Tel: +61 2 9428 6555 Fax: +61 2 9428 6352

ECREP DE Cochlear Deutschland GmbH & Co. KG

Mailänder Straße 4 a, 30539 Hannover, Germany Tel: +49 511 542 770 Fax: +49 511 542 7770

CHREP CH Cochlear AG

Peter Merian-Weg 4, 4052 Basel, Switzerland Tel: +41 61 205 8204 Fax: +41 61 205 8205

US Cochlear Americas

10350 Park Meadows Drive, Lone Tree, CO 80124, USA Tel: +1 303 790 9010

CA Cochlear Canada Inc

2500-120 Adelaide Street West, Toronto, ON M5H 1T1, Canada

Tel: +1 (800) 483 3123 Fax: +1 416 972 5083

GB UK Responsible Person: Cochlear Europe Ltd

6 Dashwood Lang Road, Bourne Business Park, Addlestone, Surrey KT15 2HJ, United Kingdom

Tel: +44 1932 26 3400 Fax: +44 1932 26 3426

BE Cochlear Benelux NV

Schaliënhoevedreef 20 i, B-2800 Mechelen, Belgium Tel: +32 15 79 55 11 Fax: +32 15 79 55 70

FR Cochlear France S.A.S.

135 Route de Saint-Simon, 31035 Toulouse, France Tel: +33 5 34 63 85 85 (International) or 0805 200 016 (National)

Fax: +33 5 34 63 85 80

IT Cochlear Italia S.r.l.

Via Trattati Comunitari Europei 1957-2007 n.17, 40127 Bologna (BO), Italy

Tel: +39 051 601 53 11 Fax: +39 051 39 20 62

SE Cochlear Nordic AB

Konstruktionsvägen 14, 435 33 Mölnlycke, Sweden Tel +46 31 335 14 61 Fax +46 31 335 14 60

www.cochlear.com

TR Cochlear Tıbbi Cihazlar ve Sağlık Hizmetleri Ltd. Şti.

Küçükbakkalköy Mah, Defne Sok, Büyükhanlı Plaza No:3 Kat:3 Daire: 9-10-11-12, 34750, Ataşehir, İstanbul, Türkiye Tel: +90 216 538 5900 Fax: +90 216 538 5919

HK Cochlear (HK) Limited

Room 1404-1406, 14/F, Leighton Centre, 77 Leighton Road, Causeway Bay, Hong Kong

Tel: +852 2530 5773 Fax: +852 2530 5183

KR Cochlear Korea Ltd

2nd Floor, Yongsan Centreville Asterium, 25, Hangang-daero 30 gil, Yongsan-gu, Seoul, Korea (04386) Tel: +82 2 533 4450 Fax: +82 2 533 8408

CN Cochlear Medical Device (Beijing) Co., Ltd

Unit 2608-2617, 26th Floor, No.9 Building, No.91 Jianguo Road.

Chaoyang District, Beijing 100022, P.R. China Tel: +86 10 5909 7800 Fax: +86 10 5909 7900

IN Cochlear Medical Device Company India Pvt. Ltd.

Ground Floor, Platina Building, Plot No C-59, G-Block, Bandra Kurla Complex, Bandra (E), Mumbai – 400 051, India Tel: +91 22 6112 1111 Fax: +91 22 6112 1100

JP 株式会社日本コクレア(Nihon Cochlear Co Ltd)

〒113-0033 東京都文京区本郷2-3-7 お茶の水元町ビル Tel: +81 3 3817 0241 Fax: +81 3 3817 0245

AE Cochlear Middle East FZ-LLC

Dubai Healthcare City, Al Razi Building 64, Block A, Ground Floor, Offices IR1 and IR2, Dubai, United Arab Emirates Tel: +971 4 818 4400 Fax: +971 4 361 8925

PA Cochlear Latinoamérica S.A.

International Business Park, Building 3835, Office 403, Panama Pacifico, Panama Tel: +507 830 6220 Fax: +507 830 6218

NZ Cochlear NZ Limited

Level 4, Takapuna Towers, 19-21 Como St, Takapuna, Auckland 0622, New Zealand Tel: + 64 9 914 1983 Fax: 0800 886 036

This material is intended for health professionals. If you are a consumer, please seek advice from your health professional about treatments for hearing loss. Outcomes may vary, and your health professional will advise you about the factors which could affect your outcome. Always read the instructions for use. Not all products are available in all countries. Please contact your local Cochlear representative for product information.

Cochlear, 科利耳, コクレア, 코클리어, Hear now. And always, Osia, SmartSound, the elliptical logo, and marks bearing an ® or ™ symbol, are either trademarks or registered trademarks of the Cochlear group of companies (unless otherwise noted).

© Cochlear Limited 2023. All rights reserved. 2023-03